NSN 6630-01-187-2930

Alcohol Analyzer - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6630-01-187-2930

Analysis Cell Type:

Standard

Special Features:

One to nine copies of test results; up to three digit sample accuracy; internal or external standard capability; sophisticated print-out (dept name, operators name, subjects name, etc.); features manual sample take; ir purge for running standard only; retest; clear test; records datd; new test sequence; enters data shown in display; infrared source is nichrome helix around ceramic core; source energy modulated; sample cell heated to 45 deg. C.

Frequency In Hertz:

180.0 or 300.0

Shelf Life:

N/a

Unit Of Measure:

Demilitarization:

No

Fiig:

T398-d