

View Online at <https://aerobasegroup.com/nsn/6630-01-314-4292>

**Current Type:**

Ac

**Frequency Rating:**

Between 50.0 hertz and 60.0 hertz

**Product Name:**

Analyzer blood gas volumetric 115/230v 50/60hz single phase auto

**Voltage In Volts:**

115.0 and 230.0

**Special Features:**

Automatic; c/o electrodes, reagents, initial calibration and supplies necessary for start up operation; calculates and measures ph, po2 and base excess; bicarbonate and O2 saturation; permits the introduction of samples of venous blood, arterial blood and expired gas

**Supplementary Features:**

Xcentral procurement of medical equipment "cenprome" service controlled and managed item; requisition using published customer order list (col) only in accordance w/service or agency instructions; service points of contact for information and requisitioning instructions are: u.S. Army medical materiel agency, usamma/sgmma-om, ft. Detrick, md 21702-5001, av 343-7403 commercial (301) 619-7403; air force medical logistics office, afmlo/fom, ft. Detrick md 21702-5001, av 343-7487 commercial (301) 619-7487; naval medical materiel support command, navmedmat suppcom, ft. Detrick md 21702-5001, av 343-7117 commercial (301) 619-7117; requisitioner should specify the voltage and frequency required; 115 vac, 60 hz, 115 vac, 50 hz; 230 vac, 50 hz, 230 vac, 60 hz

**Specification Data:**

51404-dla120-92-d-8084 government specification

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

A503u0