

View Online at <https://aerobasegroup.com/nsn/6630-01-314-6669>

**Circuit Current For Which Designed:**

Ac

**Frequency Rating:**

Between 50.0 hertz and 60.0 hertz

**Furnished Items And Quantity:**

Electrode holder-1; support rod-1; combination laboratory grade ph electrode w/sufficient gel-1

**Indicator Type:**

Dial w/pointer

**Installation Design:**

Bench

**Smallest Graduation Unit:**

20.00 millivolts and 0.20 ph

**Measurement Range:**

+0.0/+14.0 ph single indicator 2nd range

**Scale Selection Method:**

Switch

**Accuracy Rating:**

Porm 0.1 ph unit all indicators all ranges

**Operating Voltage Rating In Volts:**

Between 120.0 ac and 230.0 ac

**Exterior Container Weight:**

40.000 pounds

**Exterior Container Cubic Measure:**

0.4836 cubic feet

**Exterior Container Quantity:**

4

**Phase:**

Single

**Special Features:**

In reusable container; for measuring hydrogen ion content in buffets' and media preparation

**Nuclear Hardness Critical Feature:**

Nonhardened

**Style Designator:**

Rectangular surface mtg

**Unit Of Issue Cube:**

0.1209 cubic feet

**Unit Of Issue Weight:**

10.000 pounds

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fig:**

A310a0