

View Online at <https://aerobasegroup.com/nsn/6630-01-422-2097>

**End Application:**

For use with hand held analyzers p/n 210000 nsn 6630-01-411-2405 and p/n 111000 nsn 6630-01-411-2568.

**Product Name:**

Simulator quality control test clinical analyzer

**Special Features:**

The simulator produces signals at two levels beyond the analyzer's reportable ranges to check both the accuracy and sensitivity of the electrical measurement circuitry and the electrical isolation between individual channels. Used once a day prior to start-up. Pass/fail system readout.

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No