NSN 6630-01-643-4163

Fiig: T01200

Chemical Analysis Equipment Test Set Subassembly - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6630-01-643-4163 **Mounting Method:** Clamp **Major Components:** Seal, 2 o-rings, springs duckbill **End Application:** Gas chromograph **Component Quantity: Product Name:** Spme microseal 21-01 **Special Features:** Flurocarbon elastomer-synthetic rubber **Functional Description:** Seals the port during syringe injection and during analysis for gas chromatographysuppguse only 23 guage spme probes, may not seal properly or can cause damage to seal **Supplementary Features:** Use only 23 guage spme probes, may not seal properly or can cause damage to seal Shelf Life: N/a **Unit Of Measure: Demilitarization:** No