## NSN 6635-00-591-1656

Material Bursting Point Tester - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6635-00-591-1656

| Carrying Case:   |
|--|
| Provided   |
| Functional Classification:   |
| Aa-8.14  |
| Special Features:  |
| 7.250 in.Lg, 2.000 in.W; checks usaf grade a doped fabric; portable; pressure range 9.5 lbs; 5 tests rqmt-unsuitable fabric fails in excess of |
| 50 pct; test conditions 70 degrees f & 65 pct relative humidity; cognizant service: air force  |
| Functional Description:  |
| Small portable tester used to check usaf grade a doped fabric  |
| Entry Date:  |
| 70-04-01   |
| Reference Data And Literature:   |
| To 33a8-4-2-11 & 33a8-4-2-14   |
| Relationship To Similar Equipment:   |
| A bench type tester 6635-00-526-9423 is used when fabric is removed from acft  |
| Type/model Designation:  |
| Tt-20/e  |
| Shelf Life:  |
| N/a  |
| Unit Of Measure:   |
|  |
| Demilitarization:  |
| No   |
| Fiig:  |
| T134-m   |
|  |
|  |
|  |
|  |