NSN 6635-00-641-3483

Concrete Beam Flexural Strength Tester - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6635-00-641-3483

General Description:

16 in. By 6 in. Cross-sectional size of concrete beam specimen accommodated; loading unit data, 3rd point type loading unit, hand screw jack operation, proving ring and dial indicator measuring device; capacity data, 1250 psi maximum flexural strength capacity, 15000 lb maximum load capacity; w/tow handle, 2 wheels; w/carrying case for gage bar or proving ring and dial indicator; furnished w/5 ea steel cased molds for casting beam specimens /collapsible/, 21 in. Lg; dim. Data, h, a/a, w, a/a; no special features

Departure From Cited Document:

Αs	mod	by	mil-t-	17797	tow	handle
----	-----	----	--------	-------	-----	--------

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

A23900