NSN 6635-00-935-6872

Concrete Beam Flexural Strength Tester - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6635-00-935-6872

General Description:

6 in. By 6 in. Cross sectional size of concrete beam specimen accommodated; 3rd point type loading unit, handscrew jack operation, proving ring and dial indicator measuring device; 1250 psi maximum flexural strength capacity, 25000 lb maximum load capacity; w/tow handle, 2 wheels; w/carrying case for proving ring and dial indicators; furnished w/molds for casing beam specimens /collapsible/, 21 in. Lg; dim. 42 in. H, 15 in. W

Departure From Cited Document:

Mil-t-17797 as modified by w/tow handle, 2 wheels

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

A23900

Mil-std (military Standard):

Mil-t-17797 spec.