

View Online at <https://aerobasegroup.com/nsn/6635-01-381-1930>

General Description:

Consists of block of metal 6 in. Lg, 2 in. W, 2 in h, 3 lbs; used as reference standard to calibrate ultrasonic test equipment for fire extinguishing container tolerances

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

A23900