NSN 6635-01-396-2798

Stationary Magnetic Inspection Unit - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6635-01-396-2798

Current Type:
Ac and dc
Furnished Items And Quantity:
Copper rod 1; aux cables 4; 100 watt black light 1; hood 1; set of curtains 1; white light source 1; ventilation fan 1; set of copper braid
pads for headstock and tailstock contact pads1; suspension tank 1; hoses with spraynozzles; agitation system 1; curcuit breaker box 1;
air pressure regulator 1; air filter
Current Rating In Amps:
5000.0 or 6000.0
Maximum Demagnetizing Current Rating In Amps:
5000.0
Err Default:
Err default
Automatic Cycle Timer:
Not included
Err Default:
Err default and err default
Voltage In Volts:
200.0 and 480.0
Phase:
Three
Special Features:
Max contact head opening 52 in. Porm 2 in.; wet horizontal machine; output: dc 6000 amps with not more than 6pct ripple, ac 5000 amps;
digital ammeter displays ac rms amps or dc amps atcoil or contacts; current continuouslyvariable; headstock pressure foot switch activated;
input air pressure 80-90psi
Frequency In Hertz:
50.0 or 60.0
Functional Description:
Capable of magnetizing workpiecesin logitudinal and circular directionsof ferromagnetic parts
Specification Data:
81348-a-a-58049 government specification
Shelf Life:
N/a
Unit Of Measure:

Demilitarization:
No
Fiig:
T264-a