NSN 6635-01-508-1519

X-ray Safety Interlock Assembly - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6635-01-508-1519 **Overall Length:** 8.000 inches **Overall Height:** 6.000 inches **Overall Width:** 7.000 inches **End Application:** 6635-01-394-5926 **Accessory Components And Quantity:** Ca1 cable ass'y for lorad x-ray unit; ca2 cable ass'y for sperry x-ray nit; ca3 interlock power cable **Voltage In Volts And Current Type:** 115.0 ac **Functional Description:** Used with lorad 160kv and sperry 300kv x-ray units. When installed the interlock ensures the x-ray warning lights are connected before permitting the x-ray tube to be energized **Reference Data And Literature:** Pd74-322 **Relationship To Similar Equipment:** REPL-AC-ES6-6350 **Surface Treatment:** Paint **Purchase Description Identification:** 98752-pd74-322 Shelf Life: N/a **Unit Of Measure:**

Demilitarization:

Yes - demil/mli