NSN 6640-00-001-9167

Laboratory Buret - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6640-00-001-9167 Scale Marking: Every tenth graduation numbered **Capacity Rating:** 10.000 milliliters **Calibration Function:** To deliver **Primary Container Type:** Box **Scale Graduation Unit:** 0.05 milliliters single interval single scale **Primary Container Unit Quantity: Graduation Marking Method:** Etched **Scale Reading Direction:** Top to bottom inside single scale **Scale Graduation Unit Range:** +0.0 to 10.0 milliliters single interval single scale **Special Features:** W/polytetrafluoroethylene plug **Style Designator:** Precision and general purpose type **Specification Or Standard:** 1 type and 1 style and a class and 2 grade **Departure From Cited Document:** As modified by furnished items and quantity Shelf Life: N/a **Unit Of Measure: Demilitarization:**

No Fiig: A24200