

View Online at <https://aerobasegroup.com/nsn/6640-00-059-7165>

**Scale Marking:**

Every tenth graduation numbered

**Capacity Rating:**

Between 4.960 milliliters and 5.040 milliliters

**Calibration Function:**

To deliver w/volume adjusted for blowing out last drop

**Primary Container Type:**

Carton

**Scale Graduation Unit:**

0.1 milliliters single interval single scale

**Primary Container Unit Quantity:**

1

**Graduation Marking Method:**

Etched or fired enamel or fired stain

**Total Primary Container Quantity:**

1

**Scale Portion Outside Diameter:**

7.50 millimeters

**Scale Reading Direction:**

Top to bottom outside single scale

**Scale Graduation Unit Range:**

+0.0 to 5.0 milliliters single interval single scale

**Product Name:**

Pipet serological borosilicate glass cotton mouth 5 ml capacity

**Special Features:**

Tempered tip; 1 blue band; constricted for cotton plugging

**Material:**

Glass, borosilicate

**Material:**

Glass, borosilicate

**Style Designator:**

Expanded mouth serological

**Specification Or Standard:**

li style

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fig:**

A24200