## NSN 6640-00-202-1704

Laboratory Buret - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6640-00-202-1704 Scale Marking: Every tenth graduation numbered **Capacity Rating:** 50.000 milliliters **Calibration Function:** To deliver **Accuracy Certifier:** National bureau of standards **Primary Container Type:** Box **Scale Graduation Unit:** 0.1 milliliters single interval single scale **Primary Container Unit Quantity: Graduation Marking Method:** Etched or fired stain **Scale Reading Direction:** Top to bottom outside single scale **Scale Graduation Unit Range:** +0.0 to 50.0 milliliters single interval single scale Material: Glass, federal dd-g-541, type 1, class a **Style Designator:** Precision and general purpose type **Specification Or Standard:** 1 type and 1 style and b class and 6 size Shelf Life: N/a **Unit Of Measure: Demilitarization:** No Fiig:

A24200