NSN 6640-00-290-6837

Laboratory Buret - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6640-00-290-6837

Scale Marking:

Every tenth graduation numbered

Capacity Rating:

25.000 milliliters

Calibration Function:

To deliver

Accuracy Certifier:

National bureau of standards

Primary Container Type:

Box

Scale Graduation Unit:

0.1 milliliters single interval single scale

Primary Container Unit Quantity:

1

Graduation Marking Method:

Etched or fired stain

Scale Reading Direction:

Top to bottom outside single scale

Scale Graduation Unit Range:

+0.0 to 25.0 milliliters single interval single scale

Material:

Glass, federal dd-g-541, type 1, class a

Style Designator:

Precision and general purpose type

Specification Or Standard:

1 type and 1 style and b class and 5 size

Shelf Life:

N/a

Unit Of Measure:

Demilitarization:

No

Fiig:

A24200