## NSN 6640-00-359-9628

Laboratory Foam Test Apparatus - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6640-00-359-9628

Unit	Ot	Issue	Cube:

18.000 cubic feet

## **General Description:**

C/oborosilicate glass jar w/stirrer, temp range 75-200 degrees f, 12 in. X 18 in.- cover w/fitted 1000 ml graduated cylinder in center- diffusion stone assembly- flowmeter w/needle valve to maintain flow rate of 94 ml/min- rubber stopper- gravity thermometer m5 to p215 degrees f- 115 or 230v, 50 or 60 hz

Sha	Ιf	Life:
SHE	"	LIIE.

N/a

**Unit Of Measure:** 

--

Demilitarization:

No