

View Online at <https://aerobasegroup.com/nsn/6640-00-437-3000>

**Scale Marking:**

Every tenth graduation numbered

**Capacity Rating:**

1.000 milliliters

**Calibration Function:**

To deliver w/volume adjusted for blowing out last drop

**Primary Container Type:**

Carton

**Scale Graduation Unit:**

0.01 milliliters single interval single scale

**Primary Container Unit Quantity:**

1

**Graduation Marking Method:**

Etched or fired enamel or fired stain

**Total Primary Container Quantity:**

1

**Scale Portion Outside Diameter:**

6.00 millimeters

**Scale Reading Direction:**

Top to bottom outside single scale

**Scale Graduation Unit Range:**

+0.0 to 1.0 milliliters single interval single scale

**Product Name:**

Pipet serological borosilicate glass 1 ml

**Special Features:**

Tempered tip, 1 yellow band

**Material:**

Glass, borosilicate

**Material:**

Glass, borosilicate

**Style Designator:**

General purpose serological

**Specification Or Standard:**

I style

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fig:**

A24200