

View Online at <https://aerobasegroup.com/nsn/6640-00-437-4000>

**Scale Marking:**

Every tenth graduation numbered

**Capacity Rating:**

5.000 milliliters

**Calibration Function:**

To deliver w/volume adjusted for blowing out last drop

**Primary Container Type:**

Carton

**Scale Graduation Unit:**

0.1 milliliters single interval single scale

**Primary Container Unit Quantity:**

1

**Graduation Marking Method:**

Etched or fired enamel or fired stain

**Total Primary Container Quantity:**

1

**Scale Portion Outside Diameter:**

7.50 millimeters

**Scale Reading Direction:**

Top to bottom outside single scale

**Scale Graduation Unit Range:**

+0.0 to 5.0 milliliters single interval single scale

**Product Name:**

Pipet serological borosilicate glass 5ml

**Special Features:**

Tempered tip, 1 blue band

**Material:**

Glass, borosilicate

**Material:**

Glass, borosilicate

**Style Designator:**

Reduced mouth

**Specification Or Standard:**

li style

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fig:**

A24200