## NSN 6640-00-437-4000

Serological Pipet - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6640-00-437-4000 Scale Marking: Every tenth graduation numbered **Capacity Rating:** 5.000 milliliters **Calibration Function:** To deliver w/volume adjusted for blowing out last drop **Primary Container Type:** Carton **Scale Graduation Unit:** 0.1 milliliters single interval single scale **Primary Container Unit Quantity: Graduation Marking Method:** Etched or fired enamel or fired stain **Total Primary Container Quantity: Scale Portion Outside Diameter:** 7.50 millimeters **Scale Reading Direction:** Top to bottom outside single scale **Scale Graduation Unit Range:** +0.0 to 5.0 milliliters single interval single scale **Product Name:** Pipet serological borosilicate glass 5ml **Special Features:** Tempered tip, 1 blue band Material: Glass, borosilicate Material: Glass, borosilicate **Style Designator:** Reduced mouth **Specification Or Standard:** li style Shelf Life: N/a **Unit Of Measure: Demilitarization:** 

No

Fiig:

A24200