

View Online at <https://aerobasegroup.com/nsn/6640-00-515-5034>

Scale Marking:

Every tenth graduation numbered

Capacity Rating:

Between 1.980 milliliters and 2.020 milliliters

Calibration Function:

To contain

Scale Graduation Unit:

0.1 milliliters single interval single scale

Graduation Marking Method:

Engraved or etched or fired enamel or fired stain

Scale Portion Outside Diameter:

6.75 millimeters

Scale Reading Direction:

Bottom to top inside single scale

Scale Graduation Unit Range:

+0.0 to 2.0 milliliters single interval single scale

Style Designator:

General purpose serological

Specification Or Standard:

6 type and b class and a composition

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

A24200