

View Online at <https://aerobasegroup.com/nsn/6640-00-879-4076>

Design Designation:

Mohr

Scale Marking:

Every tenth

Capacity Rating:

2.000 milliliters

Calibration Function:

To contain

Scale Graduation Unit:

0.1 milliliters single interval single scale

Graduation Marking Method:

Engraved or etched or fired enamel or fired stain

Scale Portion Outside Diameter:

7.00 millimeters

Scale Reading Direction:

Bottom to top inside single scale

Scale Graduation Unit Range:

+0.0 to 2.0 milliliters single interval single scale

Style Designator:

Plain mouth

Specification Or Standard:

4 type and a class and b composition

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fig:

A24200