NSN 6640-01-068-0313

Laboratory Glass Sealing Machine - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6640-01-068-0313

General Description:

Fuel requirement be	ottled propane or na	atural gas; ox	kygen requirem	ent commercial gi	rade w/regulato	or to deliver 20 psi;	sealing rate	20
ampules in 5 min;	110 vac, 50/60 hz;	1 20 ampule	carousel rack;	adjustable flame	height, for 2, 5,	10 or 20 ml ampu	es; 8 in. W,	16.5 in. L,
16 in. H								

N/a			
Unit	Of	Measure:	

Demilitarization:

Shelf Life:

No