

View Online at <https://aerobasegroup.com/nsn/6640-01-082-7775>

Design Designation:

V.D.R.L.

Scale Marking:

Every graduation numbered

Capacity Rating:

Between 0.980 milliliters and 1.020 milliliters

Calibration Function:

To contain

Scale Graduation Unit:

0.1 milliliters single interval single scale

Graduation Marking Method:

Fused pigment into glass

Scale Portion Outside Diameter:

7.00 millimeters

Scale Reading Direction:

Bottom to top inside single scale

Scale Graduation Unit Range:

+0.0/+1.0 milliliters single interval single scale

Material:

Glass, federal dd-g-541, type 1, class a

Style Designator:

General purpose serological

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fig:

A24200