

View Online at <https://aerobasegroup.com/nsn/6640-01-360-5719>

**Component Quantity:**

114

**Nonsupply Items And Quantities:**

N475ml btl titration solution 1; 13ml sodium hydroxide titration cartridge 2; vial of reference control a 3; vial of reference control b 3; vial of titration acid 5; titration cups 50; sheet of 50 pipettor labels 1

**Product Name:**

Calibration kit pipet volumetric sufficient to perform 50 tests

**Special Features:**

Jw/o dispenser; contains sufficient quantity to perform 50 pipet volume verification tests performed with streck vc 100 acid base titration system; verifies calibration of any pipettor from 1 to 1000 microliters

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

A23800