

View Online at <https://aerobasegroup.com/nsn/6640-01-416-7345>

Overall Height:

10.000 inches

Current Type:

Ac

Frequency Rating:

50.0 hertz

Overall Diameter:

8.000 inches

Accessory Components And Quantity:

12 place rotor which can accomodate from 10 x 75 mm to 12 x 75 mm tubes (not provided).

Universal Medical Device Nomenclature System Code:

10780

Operation Method:

Electric

Maximum Speed Rating In Rpm:

3400.0

Guard Bowl:

Included

Furnished Controls:

High/low speed switch; electric timer; brake switch; lid safety switch.

Product Name:

Centrifuge laboratory 3400 rpm 110/220 v 50hz ac single phase

Voltage In Volts:

110.0 and 220.0

Phase:

Single

Special Features:

Portable; for blood grouping, cross matching, coombs test, and other cell washing procedures.

Supplementary Features:

Note to requisitioner: for replacement 12 place rotor order nsn 6640-00-445-1609 and for replacement 6 place rotor order nsn 6640-01-162-2949.

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

T149-c