

View Online at <https://aerobasegroup.com/nsn/6640-01-439-2807>

Frequency Rating:

60.0 hertz

End Application:

6665-00-682-4765 water test kit, bacteriological

Universal Medical Device Nomenclature System Code:

15151

Voltage Rating:

115.0 volts

Operating Temperature Range In Deg Celsius:

+30.0/+44.5

Product Name:

Incubator bacteriol rechargeable nickel-cadmium battery 115v

Phase:

Single

Special Features:

Single chamber; includes: rechargeable nickel-cadium battery, battery charger, external power supply, adapter cables, thermometer, and carrying case; 115 v; lightweight; for field use

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fig:

T269-e