## NSN 6640-01-533-2171

Laboratory Blender - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6640-01-533-2171

Manus Com Tomas
Mounting Type:
Bench
Voltage Rating:
110.0 volts
Prime Mover Current Type:
Ac
Product Name:
Food pathogen identifier pulsifier
Special Features:
250ml sample capacity; 15-180 seconds cycle time; 460x320x320mm exterior dimensions; 8.5kg weight; prepares food samples for
analysis of pathogens employing standard bacteriological methods; agitates food at a high frequency producing a combination of shock
waves which detaches bacteria into suspension; the result is a cleaner less particulate suspension making sample collection by pipette
easier than current methods
Shelf Life:
N/a
Unit Of Measure:
Demilitarization:
No
Fiig:
T269-I