

View Online at <https://aerobasegroup.com/nsn/6640-01-643-7236>

Current Type:

Ac

Frequency Rating:

Between 50.0 hertz and 60.0 hertz

Universal Medical Device Nomenclature System Code:

15173

Voltage Rating:

Between 100.0 volts and 240.0 volts

Product Name:

Rotator laboratory device code 15173

Special Features:

Horizontal mixing of different tube/vessel types; compact design for placing in incubators (non-co2) chambers, and refrigerators (up to 60 degrees centigrade); pid control ensures consistent and smooth orbital motion; speed range: 50-330 rpm; orbital size: 13mm (0.51 in); maximum load: 8.8 lbs; exterior dimensions: 13 in l x 10.9 in w x 5.1 in h; electrical requirements: 100 to 240v, 50/60 hz; platform dimensions: 9.8 in x 9.8 in

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

T269-g