NSN 6665-01-560-9344

No Fiig: A23800

Gas Indicators And Alarms Calibration Kit - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6665-01-560-9344

Accessory Component Quantity:
3
Furnished Items:
Chillgard rt or model 3800 1; a calibration check gas, air, zero 1; a calibration check gas, 400 ppm carbon monoxide in air
End Application:
Monitor, cbn monoxide
Model Number:
3800
Shelf Life:
N/a
Unit Of Measure:
Demilitarization: