NSN 6695-01-484-1625

Primary Reference Standards Calibration Set - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6695-01-484-1625

	-			
End	Λn	nlic	atio	'n.
LIIU	Δ	DIIC	auc	"

Used on mk 41 vertical launch system

Special Features:

17 liter replacement cylinder used to calibrate a gas monitor; 240 psi working press; burst press 600 psi; 300 psi test pressure; 14 oz propane cylinder, contains 25 ppm carbon monoxide gas; Igth 10.562 in; diameter 2.938 in

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

A23800