

View Online at <https://aerobasegroup.com/nsn/6840-00-116-9364>

**Physical Form:**

Liquid

**Product Type:**

Any acceptable, except mercury compounds, formaldehyde, or peroxides

**Composition Percentage By Weight:**

Any acceptable

**Test Organism Type:**

Micrococcus pyogenes var. Aureus and salmonella schottmuelleri and trichophyton interdigitale

**Dilution By Volume For Testing:**

Dilution by volume for testing 1.530 min. And 21.100 max

**Test Time Limit:**

10.0 minutes

**Test Temp:**

20.0 degrees celsius

**Special Features:**

Concentrated; suitable for cold disinfection of surgical instruments

**Specification Or Standard:**

L type

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

T090-b