## NSN 6850-01-460-3252

Turbidity Standard Solution - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6850-01-460-3252

## **Product Name:**

Stablcal turbidity standards calibration sealed vials kt f 2100p

## **Special Features:**

Support item for the water quality analysis set; stablcal standards kt, for 2100p portable turbidimeter; usepa accepted for calibration

purposes; stablilized formazin turbidity standards are true formazin dilutions developed for use in any trubidimeter; kit includes sealed vials of each: <0.1 ntu, 20 ntu, 100 ntu, 800 ntu

Shelf Life:

N/a

Unit Of Measure:

--

## Demilitarization:

No

Fiig:

T093-a